

Multi-country validation of SAMBA - A novel molecular point-of-care test for HIV-1 detection in resource-limited setting

Running title: Multi-country SAMBA HIV-1 Qual Whole Blood Test

Johnson Ondiek¹ MSc, Zikulah Namukaya² MPH, Sekesai Mtapuri-Zinyowera³ PhD, Suna Balkan⁴ (MD), Ali Elbireer^{5,11} PhD, Ines Ushiro Lumb⁶ MSc, FRCPath, Charles Kiyaga⁷ MSc, Neha Goel⁸ MSc, Allyson Ritchie⁸ PhD, Patience Ncube³ MSc, Kenneth Omuomu¹ MSc, Kenneth Ndiege¹ MSc, Adeodata Kekitiinwa⁹ MMed, Douglas Mangwanya¹⁰ MSc, Mary Glenn Fowler¹¹ MD, MPH, Lou Nadala¹², PhD, Helen Lee^{8,12} PhD

1. Kenya Medical Research Institute/US CDC Research and Public Health Collaboration (KEMRI-CDC), Kisumu, Kenya
2. Makerere University-Johns Hopkins University Research Collaboration, Kampala, Uganda
3. National Microbiology Reference Laboratory, Southerton, P.O. Box ST749, Harare, Zimbabwe
4. Médecins Sans Frontières, Paris, France
5. African Society for Laboratory Medicine, Addis Ababa, Ethiopia
6. NHS Blood and Transplant, Charcot Road, London NW9 5BG
7. Central Public Health Laboratory, Ministry of Health, P. O. Box 2210, Kampala, Uganda
8. Diagnostics Development Unit, Department of Haematology, University of Cambridge, Chesterford Research Park, Great Chesterford CB10 1XL, UK
9. Baylor-Uganda, Paediatric Infectious Diseases Clinic, Mulago Hospital, Kampala, Uganda
10. Ministry of Health and Child Care, Causeway, P.O. Box CY1122, Harare, Zimbabwe
11. Johns Hopkins University, School of Medicine, Baltimore, MD 21205, USA
12. Diagnostics for the Real World Ltd., Sunnyvale, CA 94085, USA

Correspondence to: Helen Lee, Diagnostics Development Unit, Department of Haematology, University of Cambridge, Science Village, Chesterford Research Park, Great Chesterford CB10 1XL, UK. Tel.: +44-1223-785335.

Email: hl207@cam.ac.uk

Conflicts of Interest and Source of Funding: This study was supported by a grant from the Children Investment Fund Foundation (CIFF) and UNITAID.

N.G. and A.R. are consultant and holds equity in Diagnostics for the Real World Ltd. (DRW), a spin-off company from the University of Cambridge that markets the SAMBA technology developed at the university. HL is a founder and equity holder of Diagnostics for the Real World Ltd. The University of Cambridge is also an equity holder of DRW. The remaining authors declare no conflicts of interest.

Word count
Abstract 212
Text 3,394

Abstract

Introduction: Early diagnosis of HIV-1 infection and the prompt initiation of antiretroviral therapy are critical to achieving a reduction in the morbidity and mortality of infected infants. The SAMBA HIV-1 Qual Whole Blood Test was developed specifically for early infant diagnosis and prevention of mother-to-child transmission programs implemented at the point-of-care in resource-limited settings.

Methods: We have evaluated the performance of this test run on the SAMBA I semi-automated platform with fresh whole blood specimens collected from 202 adults and 745 infants in Kenya, Uganda, and Zimbabwe. Results were compared with those obtained with the Roche COBAS AmpliPrep/COBAS TaqMan (CAP/CTM) HIV-1 assay as performed with fresh whole blood or dried blood spots of the same subjects, and discrepancies were resolved with alternative assays.

Results: The performance of the SAMBA and CAP/CTM assays evaluated at five laboratories in the three countries was similar for both adult and infant

samples. The clinical sensitivity, specificity, and positive and negative predictive values for the SAMBA test were 100%, 99.2%, 98.7%, and 100%, respectively, with adult samples, and 98.5%, 99.8%, 99.7%, and 98.8%, respectively, with infant samples.

Discussion: Our data suggest that the SAMBA HIV-1 Qual Whole Blood Test would be effective for early diagnosis of HIV-1 infection in infants at point-of-care settings in sub-Saharan Africa.

Key words: HIV-1, point-of-care, early infant diagnosis, nucleic acid test

Introduction

The UNAIDS 2014 Gap Report estimated that there were 3.2 million children infected with HIV worldwide, 91% of whom were living in sub-Saharan Africa.¹ New infections in children had fallen by 50% between 2010 and 2015,² indicating that substantial progress is being made in the prevention of mother-to-child transmission. In key high-burden countries in Africa, however, early infant diagnosis (EID) coverage was estimated to range from as high as 42% in Kenya to as low as 4% in Chad and Nigeria, so that challenges remain to the goal of reducing childhood morbidity and mortality.¹ Early diagnosis of HIV-1 infection in infants followed by prompt initiation of antiretroviral therapy (ART) is critical to prevent progression of disease and to reduce mortality, given that about one-third of infected children die by their first birthday and half die by their second.³ WHO guidelines therefore recommend that ART be initiated without delay in infants with an initial positive virological test result.⁴

In contrast to adults, for whom rapid antibody tests are readily available and affordable, accurate serological HIV-1 diagnosis in children younger than 18 months is not possible because of the passive transfer of maternal antibodies during pregnancy. Diagnosis of HIV-1 infection in infants therefore requires sophisticated and specialized laboratory diagnostic tests to detect viral nucleic acid or viral products such as the p24 antigen.^{5 6 7 8} The expense, complexity, and infrastructure requirements of these technologies mean that access to timely testing services is highly limited for HIV-exposed infants in many resource-limited settings. Low EID rates contribute substantially to the poor or nonexistent status of links between prevention of mother-to-child

transmission and pediatric treatment programs, with only 49% of HIV-infected infants in low- and middle-income countries worldwide receiving treatment by the end of 2015.² The current practice for EID in many resource-limited settings is based on the collection of dried blood spot (DBS) samples on filter paper from infants attending peripheral clinics and the forwarding of these specimens to centralized laboratories for molecular testing to detect HIV-1 nucleic acid. Although testing of samples may take only a few hours to a day, the time taken in their collection, storage, transportation, processing and forwarding of results can result in delays of up to 2 months before results are provided to the patient.⁹ As a result of this long turnaround time, loss to follow-up has become one of the major problems in the provision of effective pediatric treatment and care for HIV-1–infected children.¹⁰

Point-of-care (POC) nucleic acid tests could shorten the turnaround time for test results from weeks to hours and thereby greatly reduce loss to follow-up. The utility of POC tests will depend on their simplicity of use, affordability, and stability at room temperature as well as on their ability to deliver results in field conditions with an accuracy and precision similar to those for standard quantitative or qualitative diagnostic tests performed in centralized facilities.^{11 12 13 14}

The SAMBA (Simple AMplification-Based Assay) HIV-1 Qual Whole Blood Test was designed for the qualitative detection of both HIV-1 proviral DNA and RNA in whole blood of infected adults and infants, with the result provided via a visual readout on a dipstick.¹⁵ The test was also specifically designed for use in POC settings, having a total assay time of approximately 2 hours. We now present the results of a cross-sectional field evaluation of the SAMBA HIV-1 Qual Whole Blood Test performed with clinical samples obtained from adults and infants in three countries in sub-Saharan Africa with a high prevalence of HIV-1 infection.

Methods

Research ethics and informed consent

The study was performed in accordance with the Declaration of Helsinki and participating countries research and ethical regulations. Ethical approval was obtained from the Scientific Steering and Ethical Review Committee of the

Kenya Medical Research Institute–Center for Disease Control (KEMRI-CDC) in Kisumu, Kenya; from the Uganda National Council for Science and Technology for Arua District Hospital, Uganda; from the Joint Clinical Research Centre Ethics Committee, the Uganda National Council for Science and Technology, and the Institutional Review Board of Baylor College of Medicine for the Makerere University–Johns Hopkins University (MU-JHU) Research Collaboration and Baylor College of Medicine Children’s Foundation (Baylor-Uganda), Kampala; and from the Institutional Review Board of Harare Central Hospital and the Medical Research Council of Zimbabwe for the National Microbiology Reference Laboratory, Harare, Zimbabwe. All adult subjects provided written informed consent for study participation, with the parents or guardians of infant subjects also providing such consent.

Sample collection and preparation

Whole blood was collected either by venepuncture into BD Vacutainer K2-EDTA tubes (Becton Dickinson, Franklin Lakes, NJ) or, in the case of Kenyan infants, by heel or finger pricks. Specimens from adults were initially tested for the presence of antibodies to HIV-1 at the health care setting as per the standard of care. In order to limit patient burden on sample collection venepuncture sample collection was carried out. EDTA samples collected for CD4 count were used in some cases. Laboratory technicians prepared DBS specimens in the testing laboratory by pipetting 70 µl of K2-EDTA whole blood onto each 4 circles of two Whatman 903 cards, which were then dried overnight and packed into individual zip lock foil bags with desiccant. One card was maintained at ambient temperature before transportation to the testing laboratories within 72 hours or was stored at –20°C for up to two weeks until transfer. The second card was retained frozen for quality control purposes or discrepant analysis. At sites using the London laboratories for discrepant analysis, two additional 200-µl samples of whole blood were frozen onsite at –20°C and subsequently shipped on dry ice to the Virology laboratories at the Royal London Hospital or Public Health England, both in London, UK.

Evaluation protocol and recruitment

In Kenya, 335 infants, including 135 exposed and 200 known HIV-1 positive, were recruited to the study between November 2013 and April 2014 (Table 1). The infants were recruited from support and health facilities for HIV-infected patients in Kisumu, Siaya, Kisumu, Migori, Nyamira, and Homa Bay. Whole blood samples (150µl) were tested within 24 hours of collection both with the SAMBA HIV-1 Qual Whole Blood Test (Diagnostics for the Real World) and with the Roche COBAS AmpliPrep/COBAS TaqMan (CAP/CTM) HIV-1 Qualitative test as performed by local trained technicians at KEMRI-CDC. Discrepant results were resolved by testing of whole blood specimens with the Abbott HIV-1 RealTime Assay as performed on the Abbott m2000 system, also at KEMRI-CDC (Table 1).

In Arua, Uganda, whole blood and DBS samples were collected from 102 adults recruited from Médecins sans Frontières (MSF) programs at Arua Regional Referral Hospital in September 2011. The subjects included 22 ART-naïve, HIV-1-positive patients and 80 individuals of previously unknown HIV status from the voluntary testing and counselling clinic. Whole blood samples were tested with the SAMBA assay as performed by a trained MSF technician onsite within 4 h of collection. DBS samples were sent to the Central Public Health Laboratory (CPHL) in Kampala for testing with the CAP/CTM assay. Discrepant analysis was performed at the Royal London Hospital using at the Royal London Hospital, using a HIV-1 proviral DNA PCR targeted against four different regions of the viral genome (Table 1).

In Kampala, Uganda, whole blood and DBS specimens were collected between January and September 2014 from a total of 311 infants, including 201 vertically exposed infants from MU-JHU or the postnatal clinic at Mulago Hospital and 110 known HIV-1 positive infants recruited from the pediatric infectious disease clinic at Baylor-Uganda, the Mildmay Uganda Clinic or the postnatal clinic at Mulago Hospital. These infants were also enrolled in an ongoing clinical research study. Whole blood samples were tested with the SAMBA assay at the Mulago Core Laboratory by local trained technicians within 1 - 2 hours of collection. DBS samples were sent to CPHL within 3 days of preparation for testing with the CAP/CTM assay. Frozen whole blood for all

discordant samples and 10% of randomly selected concordant samples was sent to Public Health England for testing with the four-gene proviral DNA assay (Table 1).

In Zimbabwe, whole blood and DBS samples were collected from 100 adult patients of previously unknown HIV status and from 99 exposed infants recruited from Harare Central Hospital between July and August 2014. Whole blood and DBS samples were tested within 6 hours of collection with the SAMBA and CAP/CTM assays, respectively, as performed by local trained technicians at the National Microbiology Reference Laboratory (NMRL), Ministry of Health and Child Care, in Harare. Discrepant results were resolved by testing of DBS samples with the Abbott HIV-1 RealTime Assay as performed on the Abbott m2000 system at the Cimas Private Laboratory of the Cimas Medical Aid Society in Harare (Table 1).

The SAMBA HIV-1 Qual Whole Blood Test was performed on the semi-automated SAMBA I system, consisting of SAMBAprep and SAMBAamp instruments,¹⁶ at all sites with the exception of Arua, Uganda, where a semi-automated SAMBA system consisting of prototype sample preparation and amplification-detection instruments was used. Sample preparation takes 45 minutes and amplification 60 minutes. The limit of detection is 400 copies/ml. Results were read by the user via a visual readout on a lateral flow test strip,¹⁵ ¹⁶ ¹⁷ with the presence of HIV and control lines indicating a positive result, a control line only indicating a negative result, and no lines indicating an invalid result.

All testing with the CAP/CTM and Abbott assays was performed according to the manufacturers' instructions. The four-gene proviral DNA assay was performed according to a modified protocol described previously.¹⁸ ¹⁹ Samples consisting of 200 or 350 µl of whole blood were subjected to extraction manually with the use of either the QIAamp DNA Blood Mini Kit (Qiagen, Crawley, UK) or the Qiagen EZ1 automated extraction system (EZ1 DNA program), respectively. In addition to the discordant samples, 10% to 20% (depending on the site) of the total concordant positive or negative samples were tested in a blinded manner in order to prevent bias.

All SAMBA and CAP/CTM testing was also performed in a blinded manner, and results for the two assays were then compared to determine concordance. Tie-breaker results for discrepant samples and results for randomly selected concordant positive or negative samples were also sent to the study coordinator to determine the true status of the discordant specimens. No SAMBA test results were shared with parents or guardians of infant subjects. However, if a final resolved result differed from the previously determined HIV-1 status of the subject at baseline, the result was shared with the clinicians for follow-up.

Statistical analysis

Sensitivity and specificity for the resolved data for the SAMBA HIV-1 Qual Whole Blood Test and CAP/CTM were calculated together with the corresponding 95% confidence intervals (CIs). The calculations were performed with the use of MedCalc software (https://www.medcalc.org/calc/diagnostic_test.php).

Results

Performance of the SAMBA HIV-1 Qual Whole Blood Test with adult samples

Sites in Uganda and Zimbabwe respectively tested 102 and 100 venous samples from adults (Table 1). In Uganda, there were 32 concordant positive and 68 concordant negative results with the SAMBA and CAP/CTM assays, whereas there were 43 concordant positive and 56 concordant negative results in Zimbabwe (Table 2). Three discrepant results, two in Uganda and one in Zimbabwe, were positive with the SAMBA test and negative by CAP/CTM. Frozen whole blood for the two discrepant samples from Uganda were sent to the Royal London Hospital together with two concordant positive samples and two concordant negative samples. The results of the concordant samples were confirmed, and the two discrepant samples were found to be positive, in agreement with the SAMBA test results. The discordant sample from Zimbabwe was confirmed negative by DBS testing with the Abbott HIV-1 RealTime Assay at Cimas in Harare, consistent with the result of the

CAP/CTM assay. The resolved sensitivity and specificity of the SAMBA HIV-1 Qual Whole Blood Test for the 202 adult samples were thus 100% (95% CI, 95.3–100) and 99.2% (95% CI, 95.6–99.9), respectively (Table 3). The resolved sensitivity and specificity of the CAP/CTM assay were 97.4% (95% CI, 90.9–99.7) and 100% (95% CI, 97.1–100), respectively.

Performance of the SAMBA HIV-1 Qual Whole Blood Test with infant samples

In Kenya, a total of 335 infants were enrolled in the study (Table 1). Infant samples were collected via heel or finger prick, as per the standard of care at the various sites, and fresh whole blood was used for testing with both SAMBA and CAP/CTM assays. There were 200 concordant positive and 131 concordant negative results for the two tests (Table 2). Whole blood for the four discordant samples, including three that were SAMBA negative but CAP/CTM positive and one that was SAMBA positive but CAP/CTM negative, were tested with the Abbott HIV-1 RealTime Assay at KEMRI-CDC. The three samples that were SAMBA negative were shown to be true negatives by the Abbott test, whereas the discordant SAMBA-positive sample was found to be a false positive. The resolved sensitivity, and specificity of the SAMBA HIV-1 Qual Whole Blood Test for this Kenyan cohort of 335 infant whole blood samples were thus 100% (95% CI, 98.2–100), 99.3% (95% CI, 95.9–99.9), 99.5% (95% CI, 97.3–99.9), and 100% (95% CI, 97.3–100), respectively (Table 4). The sensitivity and specificity of CAP/CTM for the same cohort were 100% (95% CI, 98.2–100), 97.8% (95% CI, 93.6–99.5), 98.5% (95% CI, 95.7–99.7), and 100% (95% CI, 97.2–100), respectively.

In Uganda, samples from 311 infants were tested (Table 1). There were 100 concordant positive and 203 concordant negative results for the two assays (Table 2). Five samples were positive by CAP/CTM but negative by SAMBA, and three samples were positive by SAMBA but negative by CAP/CTM. The eight discordant samples, together with 60 randomly selected concordant samples (positives and negatives), were tested for proviral DNA at Public Health England. The results from all of the randomly selected concordant samples demonstrated consistency after testing at VRD-PHE. The five samples that were negative by SAMBA but positive by CAP/CTM were

confirmed positive (SAMBA false negative), and the three samples that were positive by SAMBA but negative by CAP/CTM were also confirmed positive (CAP/CTM false negative). The resolved sensitivities of the SAMBA test with whole blood and of the CAP/CTM assay with DBS samples for these 311 Ugandan infants were thus 95.4% (95% CI, 89.5–98.5) and 97.2% (95% CI, 92.1–99.4), respectively (Table 4). The specificities were 100% for both assays.

In Zimbabwe, 99 infants were tested (Table 1). There were 23 concordant positive and 75 concordant negative results for the two assays (Table 2). The one discordant sample, which was positive by CAP/CTM but negative by SAMBA, was found to be negative by testing with the Abbott HIV-1 RealTime Assay at Cimas (CAP/CTM false positive). The resolved sensitivity and specificity, of the SAMBA test for this cohort of Zimbabwean infants were thus all 100% (Table 4). The resolved sensitivity and specificity of CAP/CTM were 100% (95% CI, 85.2–100), 98.7% (95% CI, 92.9–99.9), 95.8% (95% CI, 78.9–99.9), and 100% (95% CI, 95.2–100), respectively.

Overall, the results for the 745 infants tested in the three African countries showed the sensitivity of the SAMBA Qual Whole Blood Test to be 98.5% (95% CI, 96.5–99.5) and that for CAP/CTM to be 99.1% (95% CI, 97.4–99.8), with corresponding specificity values of 99.8% (95% CI, 98.7–99.9) and 99.0% (95% CI, 97.5–99.7), respectively (Table 4). In addition, the 202 adults tested showed a sensitivity of 100% for SAMBA and that for CAP/CTM to be 97.4% with corresponding specificity values of 99.2% and 100.0% (Table 3). Combination of the results for the infant samples with those for the 202 adult samples yielded a sensitivity of 98.8% (95% CI, 97.2–99.6) for both SAMBA and CAP/CTM and specificities of 99.6% (95% CI, 98.7–99.9) and 99.3% (95% CI, 98.1–99.8), respectively.

Discussion

There is an urgent need for highly accurate, affordable, robust, and easy-to-use HIV-1 tests that can be applied to EID in POC settings. The development of such tests with performance characteristics similar to those of PCR-based assays currently used for testing of DBS samples in centralized laboratories should increase access to testing, while the ability to provide test results at

the same clinic visit as sample collection should reduce loss to follow-up, facilitate timely initiation of treatment, contributing to an overall reduction in morbidity and mortality for infected infants.

Roche CAP/CTM as performed with DBS samples is currently the most widely used assay for EID of HIV-1 infection in sub-Saharan Africa. However, centralised system has an associated high equipment cost, requires considerable bench space, a controlled environment and frozen reagents; the assay is technically complex and it requires highly trained personnel in a centralized facility. The SAMBA I system is a considerably simpler instrument with a smaller footprint, it does not require a temperature-controlled environment, and it relies on freeze-dried reagents that can be stored at room temperature.¹⁵ It is therefore suitable for operation in small laboratories in district hospitals such as those in Chiradzulu, Malawi, and Arua, Uganda, where it has been in routine use since September 2013.

In the present study, the SAMBA HIV-1 Qual Whole Blood Test showed a sensitivity and specificity similar to those of the CAP/CTM assay,⁵ the current gold standard for EID in resource-limited settings. However, CAP/CTM can only be used in few centralized laboratories, and thus incurs the delays and attrition that preclude a truly effective EID program. The two assays performed similarly at sites in Kenya, Uganda, and Zimbabwe, suggesting that the performance of the SAMBA test is unaffected by differences in HIV-1 subtype frequency (predominantly subtype A in Kenya,²⁰ ²¹ subtypes A and D in Uganda,²¹ and subtype C in Zimbabwe²¹) or in geographical settings. A previous study also indicated that the SAMBA assay is highly effective with a broad coverage HIV-1 subtypes, including circulating recombinant forms.¹⁵

Our field evaluations showed that the SAMBA HIV-1 Qual Whole Blood Test performed similarly to CAP/CTM when adult and infant data were examined separately or combined. With adult or infant samples from the same individuals, the assay sensitivity was >98% for SAMBA and >97% for CAP/CTM and the specificity was at least 99% for both assays. The deficit in sensitivity of both assays was essentially due to 8 discordant infant and 2 discordant adult samples in Uganda (5 for CAP/CTM and 5 for SAMBA) that were confirmed positive in the discrepant analysis (Table 2). Unfortunately,

sample volume collected was not sufficient to allow discrepant testing with both the Abbott and proviral DNA assays.

There are some limitations to the present study. First, the median age of infants was 6.9 months at Mulago Hospital and Mildmay Uganda Clinic in Uganda, 8.7 months in Kenya, and 12 months in Zimbabwe. Given that current WHO guidelines stipulate that EID should be performed at 4 to 6 weeks of age.⁴ it will be important to determine the performance of the SAMBA HIV-1 Qual Whole Blood Test in younger infants. Second, the DBS samples were prepared for the present study by pipetting anticoagulated venous blood on the filter paper; providing samples for the CAP/CTM that were possibly of better quality than usually possible to obtain in practice. In optimal conditions, which are not always encountered in the field. The SAMBA I instrument system is a near POC designed for use in resource limited settings, such as district level hospitals and health centers, where space is restricted, throughput is low, facilities are basic, and training of staff is limited. This SAMBA-I system including two separate instruments and a minimum amount of hands on steps is adapted to the needs of small and medium size health settings processing between 16 and 48 samples/day. Another variant of the SAMBA technology – SAMBA II is a “sample-in-result out” true point of care system better adapted to next to patient, mobile clinics or even home POC. Further implementation of the SAMBA HIV-1 Qual Whole Blood Test will be undertaken in the near future.

References

1. UNAIDS. *The GAP report*. 2014.
2. UNAIDS. *Fact sheet*. 2016.
3. Newell ML, Brahmabhatt H, Ghys PD. Child mortality and HIV infection in Africa: a review. *AIDS*. 2004;18 Suppl 2:S27-34.
4. WHO. *Consolidated Guidelines on the Use of Antiretroviral Drugs for Treating and Preventing HIV Infection*. 2013.
5. Kivuyo SL, Johannessen A, Troseid M, et al. p24 antigen detection on dried blood spots is a feasible and reliable test for infant HIV infection in rural Tanzania. *Int J STD AIDS*. 2011;22(12):719-721.
6. Stevens W, Erasmus L, Moloi M, Taleng T, Sarang S. Performance of a novel human immunodeficiency virus (HIV) type 1 total nucleic acid-based real-time PCR assay using whole blood and dried blood spots

- for diagnosis of HIV in infants. *J Clin Microbiol.* 2008;46(12):3941-3945.
7. Burgard M, Blanche S, Jasseron C, et al. Performance of HIV-1 DNA or HIV-1 RNA tests for early diagnosis of perinatal HIV-1 infection during anti-retroviral prophylaxis. *J Pediatr.* 2012;160(1):60-66 e61.
 8. Panakitsuwan S, Yoshihara N, Hashimoto N, Miyamura K, Chotpitayasunondh T. Early diagnosis of vertical HIV infection in infants by rapid detection of immune complex-dissociated HIV p24 antigen. *AIDS Patient Care STDS.* 1997;11(6):429-433.
 9. Mugambi ML, Deo S, Kekitiinwa A, Kiyaga C, Singer ME. Do diagnosis delays impact receipt of test results? Evidence from the HIV early infant diagnosis program in Uganda. *PLoS One.* 2013;8(11):e78891.
 10. Phelps BR, Ahmed S, Amzel A, et al. Linkage, initiation and retention of children in the antiretroviral therapy cascade: an overview. *AIDS.* 2013;27 Suppl 2:S207-213.
 11. Jani IV, Meggi B, Mabunda N, et al. Accurate early infant HIV diagnosis in primary health clinics using a point-of-care nucleic acid test. *J Acquir Immune Defic Syndr.* 2014;67(1):e1-4.
 12. Boyle DS, Lehman DA, Lillis L, et al. Rapid detection of HIV-1 proviral DNA for early infant diagnosis using recombinase polymerase amplification. *MBio.* 2013;4(2).
 13. Jangam SR, Yamada DH, McFall SM, Kelso DM. Rapid, point-of-care extraction of human immunodeficiency virus type 1 proviral DNA from whole blood for detection by real-time PCR. *J Clin Microbiol.* 2009;47(8):2363-2368.
 14. Jangam SR, Agarwal AK, Sur K, Kelso DM. A point-of-care PCR test for HIV-1 detection in resource-limited settings. *Biosens Bioelectron.* 2013;42:69-75.
 15. Lee HH, Dineva MA, Chua YL, Ritchie AV, Ushiro-Lumb I, Wisniewski CA. Simple amplification-based assay: a nucleic acid-based point-of-care platform for HIV-1 testing. *J Infect Dis.* 2010;201 Suppl 1:S65-72.
 16. Ritchie AV, Ushiro-Lumb I, Edemaga D, et al. SAMBA HIV semiquantitative test, a new point-of-care viral-load-monitoring assay for resource-limited settings. *J Clin Microbiol.* 2014;52(9):3377-3383.
 17. Dineva MA, Candotti D, Fletcher-Brown F, Allain JP, Lee H. Simultaneous visual detection of multiple viral amplicons by dipstick assay. *J Clin Microbiol.* 2005;43(8):4015-4021.
 18. Yang C, Dash BC, Simon F, et al. Detection of diverse variants of human immunodeficiency virus-1 groups M, N, and O and simian immunodeficiency viruses from chimpanzees by using generic pol and env primer pairs. *J Infect Dis.* 2000;181(5):1791-1795.
 19. Yang C, Pieniazek D, Owen SM, et al. Detection of phylogenetically diverse human immunodeficiency virus type 1 groups M and O from plasma by using highly sensitive and specific generic primers. *J Clin Microbiol.* 1999;37(8):2581-2586.
 20. Kageha S, Lihana RW, Okoth V, et al. HIV type 1 subtype surveillance in central Kenya. *AIDS Res Hum Retroviruses.* 2012;28(2):228-231.
 21. Lihana RW, Ssemwanga D, Abimiku A, Ndembi N. Update on HIV-1 diversity in Africa: a decade in review. *AIDS Rev.* 2012;14(2):83-100.

Table 1. Sites for sample collection and testing

| Country | Sample collection site | SAMBA testing lab (sample type) | CAP/CTM testing lab (sample type) | Discrepant testing lab (test, sample type) | Age of subjects (months or years) | HIV-1 status | No. of samples |
|----------|---|---------------------------------|-----------------------------------|--|-----------------------------------|--------------|----------------|
| Kenya | Kisumu, Siaya, Kisii, Migori, Nyamira, Homa Bay | KEMRI-CDC (whole blood) | KEMRI-CDC (whole blood) | KEMRI-CDC (Abbott, whole blood) | ≤18 m | Exposed | 135 |
| | | | | | | Infected | 200 |
| Uganda | Mulago Hospital PNC | Mulago CL (whole blood) | CPHL (DBS) | PHE (proviral DNA PCR, whole blood) | ≤12 m | Exposed | 175 |
| | Baylor-Uganda PIDC | | | | | Infected | 45 |
| | MU-JHU | | | | | Infected | 29 |
| | Mildway Uganda Clinic | | | | | Exposed | 26 |
| | | | | | | Infected | 36 |
| | Arua District Hospital | MSF lab (whole blood) | CPHL (DBS) | RLH (proviral DNA PCR, whole blood) | ≥18 y | Unknown | 80 |
| | | | | | | Infected | 22 |
| Zimbabwe | Harare Central Hospital, VCT | NMRL (whole blood) | NMRL (DBS) | CIMAS-PVT (Abbott, DBS) | ≥18 y | Unknown | 100 |
| | Harare Central Hospital, pediatric ward | | | | | Exposed | 99 |

Abbreviations: SAMBA, simple amplification-based assay; CAP/CTM, Roche COBAS AmpliPrep/COBAS Taqman; KEMRI-CDC, Kenya Medical Research Institute–Center for Disease Control; MU-JHU, Makerere University–Johns Hopkins University; CL, Core Laboratory; CPHL, Central Public Health Laboratory; DBS, dried blood spot; PHE, Public Health England; PCR, polymerase chain reaction; MSF, Médecins sans Frontières; RLH, Royal London Hospital; VCT, Voluntary Counseling and Testing; NMRL, National Microbiology Reference Laboratory; CIMAS-PVT, Commercial and Industrial Medical Aid Society in Zimbabwe.

Table 2. Detailed results obtained with the SAMBA HIV-1 Qual Whole Blood test and CAP/CTM assay

| Country | Population | Concordant samples | | Discordant samples | | Confirmatory assay | | Final result |
|----------|------------|--------------------|----------|--|--|--------------------------|------------------|---------------------|
| | | Positive | Negative | SAMBA ⁺ CAP/CTM ⁻ | SAMBA ⁻ CAP/CTM ⁺ | Abbott HIV-1 RealTime | HIV-1 DNA PCR | |
| Uganda | Adult | 32 | 68 | 2 | | | Pos | 2 CAP/CTM false neg |
| Zimbabwe | Adult | 43 | 56 | 1 | | Neg | | 1 SAMBA false pos |
| Kenya | Infant | 200 | 131 | 1 | | Neg | | 1 SAMBA false pos |
| | | | | | 3 | Neg | | 3 CAP/CTM false pos |
| Uganda | Infant | 100 | 203 | 3 | | | Pos | 3 CAP/CTM false neg |
| | | | | | 5 | | Pos | 5 SAMBA false neg |
| Zimbabwe | Infant | 23 | 75 | | 1 | Neg | | 1 CAP/CTM false pos |

Table 3. Resolved performance of the SAMBA HIV-1 Qual Whole Blood Test and CAP/CTM assay with 202 adult samples

| Assay | Country | Sample type | Confirmed results | | | | Total | Sensitivity (95% CI) | Specificity (95% CI) |
|----------------|--------------|-------------|-------------------|------------|----------|----------|------------|--------------------------|--------------------------|
| | | | TP | TN | FP | FN | | | |
| SAMBA | Uganda | Venous FWB | 34 | 68 | 0 | 0 | 102 | 100% (89.7–100) | 100% (94.7–100) |
| SAMBA | Zimbabwe | Venous FWB | 43 | 56 | 1 | 0 | 100 | 100% (91.8–100) | 98.3% (90.6–99.9) |
| SAMBA | Total | | 77 | 124 | 1 | 0 | 202 | 100% (95.3–100) | 99.2% (95.6–99.9) |
| CAP/CTM | Uganda | Venous DBS | 32 | 68 | 0 | 2 | 102 | 94.1% (80.3–99.3) | 100% (94.7–100) |
| CAP/CTM | Zimbabwe | Venous DBS | 43 | 56 | 0 | 0 | 100 | 100% (91.8–100) | 100% (93.7–100) |
| CAP/CTM | Total | | 75 | 125 | 0 | 2 | 202 | 97.4% (90.9–99.7) | 100% (97.1–100) |

Abbreviations: TP, true positive; TN, true negative; FP, false positive; FN, false negative; CI, confidence interval; FWB, fresh whole blood; DBS, dried blood spot.

Table 4. Resolved performance for the SAMBA HIV-1 Qual Whole Blood Test and CAP/CTM assay with 745 infant samples

| Assay | Country | Sample type | Confirmed results | | | | Total | Sensitivity (95% CI) | Specificity (95% CI) |
|----------------|--------------|---------------|-------------------|------------|----------|----------|------------|------------------------------------|-------------------------------|
| | | | TP | TN | FP | FN | | | |
| SAMBA | Kenya | Capillary FWB | 200 | 134 | 1 | 0 | 335 | 100% (98.2–100) | 99.3% (95.9– 99.9) |
| SAMBA | Uganda | Venous FWB | 103 | 203 | 0 | 5 | 311 | 95.4% (89.5–98.5) | 100% (98.2–100) |
| SAMBA | Zimbabwe | Venous FWB | 23 | 76 | 0 | 0 | 99 | 100% (85.2–100) | 100% (95.3–100) |
| SAMBA | Total | | 326 | 413 | 1 | 5 | 745 | 98.5% (96.5–99.5) | 99.8% (98.7– 99.9) |
| CAP/CTM | Kenya | Capillary FWB | 200 | 132 | 3 | 0 | 335 | 100% (98.2–100) | 97.8% (93.6– 99.5) |
| CAP/CTM | Uganda | Venous DBS | 105 | 203 | 0 | 3 | 311 | 97.2% (92.1–99.4) | 100% (98.2–100) |
| CAP/CTM | Zimbabwe | Venous DBS | 23 | 75 | 1 | 0 | 99 | 100% (85.2–100) | 98.7% (92.9– 99.9) |
| CAP/CTM | Total | | 328 | 410 | 4 | 3 | 745 | 99.1% (97.4–99.8) | 99.0% (97.5– 99.7) |

Abbreviations: TP, true positive; TN, true negative; FP, false positive; FN, false negative; CI, confidence interval; FWB, fresh whole blood; DBS, Dried Blood Spot